

NO.3(38)-AGMC&GBPH/S&P/2024-25/Sub-I

Government of Tripura
Office of the Medical Superintendent & HOD
AGMC & GBP Hospital, Agartala


Dated, Agartala, the...../...../2025

CORRIGENDUM

The following modifications and Addendum have been done as per the pre bid meeting held on 11.02.2025 of vide tender ID-2025_OSD_57976_1 (Ref No.F.3(38)-AGMC&GBPH/S&P/2024-25/Sub-I/34377-78, dated 05.02.2025)

Here kindly read the terms & conditions - as”

1. Please read the word the **Last date of bid Submission is up to 05.00 pm of 12.03.2025 and opening of bid is 12.00 pm of 13.03.2025.**
2. Please read the word the **up-to-date GST Registration Certificate, GST clearance certificate, Professional Tax Clearance Certificate, Income Tax clearance certificate, relevant & Valid PAN Card** in place of GST Registration & up to date GST Clearance Certificate, Acknowledgement copy of income Tax Clearance Certificate & valid PAN card.
3. Please read the word **the Annual turnover of the firms /agencies for the last 03 (Three) years should not be less than 1.50 (one Crore fifty lakh)** in place of Annual turnover of the firms /agencies for the last 03 (Three) years should not be less than 3.00 (Three Crore).
4. Please read the word **Bidder have to submit Performance Certificate for last 1(one) year of the same items and which should be obtained from at least any Three (03) Govt. / State Hospital / National level Hospital / Premier Private hospitals etc** in place of Bidder have to submit last **1(one) year** Performance Certificate / Documents pertaining to supply (Supply order with execution) of the same items and should be obtained from at least any Govt. State Referral Hospital etc.
5. **Please read the specifications as**
 - i) It should be cartridge-based ISE (Ion Selective Electrode) Electrolyte analyzer.
 - ii) It should have facility for analyzing parameters like Sr. Sodium, Sr. Potassium, Sr. Chloride & Ionized Calcium.
 - iii) It should analyse samples like Whole Blood, Serum, Plasma & Urine.
 - iv) It should have Throughput of at least 80 Samples/Hour or more.
 - v) Analysis time should not be more than 40 seconds.
 - vi) Probe length should be more than 5 cm.
 - vii) It should have Touch screen display.
 - viii) It should have Bidirectional LIS.
 - ix) It should have in built Thermal Printer.
 - x) It should have facility for Internal Quality control display.
 - xi) It should have facility for Barcode scanner.
 - xii) It should have facility like display of Online LJ Graphs compliant to Westgard Rules.
 - xiii) It should have USB ports for connectivity.
 - xiv) It should have facility for Data Storage of at least 1000 samples.


4/3/25
Medical Superintendent & HOD
A.G.M.C & G.B.P. Hospital, Agartala