

GOVERNMENT OF TRIPURA
HEALTH & FAMILY WELFARE DEPARTMENT
OFFICE OF THE MEDICAL SUPERINTENDENT & HOD
AGMC & GBP HOSPITAL, AGARTALA

No.F.3(39)-AGMC&GBPH/S&P/2024-25/Sub-III/36152-53 Dated, Agartala, 18/02/2025

Short Notice Inviting Quotation (SNIQ)

An SNIQ has been invited by the Medical Superintendent & HOD, AGMC & GBP Hospital, Agartala, from reputed manufacturing firms/authorized distributors/suppliers/dealers/retailers for **procurement of certain equipment's for use at AGMC & GBP Hospital, Agartala.**

The last date of submission of quotations is up to 05.00 pm on 28.02.2025. The specifics of the SNIQ may be seen at the AGMC website www.agmc.gov.in and can also be collected in person from the office of the Medical Superintendent, AGMC & GBP Hospital, Agartala (**S&P Section**) on any working days from 20.02.2025 to 27.02.2025 between 11.00 am and 4.00 pm.

Signed by

Sankar Chakraborti

Date: 18-02-2025 13:12:21

Medical Superintendent & HOD
AGMC & GBP Hospital, Agartala

GOVERNMENT OF TRIPURA
HEALTH & FAMILY WELFARE DEPARTMENT
OFFICE OF THE MEDICAL SUPERINTENDENT & HOD
AGMC & GBP HOSPITAL, AGARTALA

No.F.3(39)-AGMC&GBPH/S&P/2024-25/Sub-III / 36152-53

Dated, Agartala. 18/02/2025

Short Notice Inviting Quotation (SNIQ)

Quotations in sealed cover are hereby invited by the Medical Superintendent & HOD, AGMC & GBP Hospital, Agartala, from reputed manufacturing firms/authorized distributors/suppliers/dealers for procurement of certain equipment's for use at AGMC & GBP Hospital, Agartala as per the following specifications.

Sl. No	Name of item(s)	Specifications	Rate to be quoted For the following units	Name of Manufacturer	Rate per Unit (Including all taxes)
(1)	(2)	(3)	(4)	(5)	(6)
1	Semi-Automated Urine chemistry analyser	As per ANNEXURE-II	Per nos	To be quoted by quotationers	To be quoted by quotationers
2	Digital Air Incubator 37 °c	Digital Temperature controller with 12"X12"X 12" in height	Per nos	To be quoted by quotationers	To be quoted by quotationers
3	Table top Centrifuge (6-8 tubes)	6-8 tubes	Per nos	To be quoted by quotationers	To be quoted by quotationers

Terms and Conditions:-

1. Quotation on plain paper addressing the Medical Superintendent & HOD, AGMC & GBP Hospital, Agartala shall reach the office of the Medical Superintendent & HOD, AGMC & GBP Hospital, Agartala, either by mail/courier or in person on or before **05.00 pm on 28.02. 2025.** Quotations shall have to be enclosed in sealed envelopes super-scribed as "**Quotation for procurement of certain equipment's for use at AGMC & GBP Hospital, Agartala, Pin-799006.**"
2. Any quotation received after the aforesaid timeline shall be liable to be summarily rejected. The Medical Superintendent & HOD, AGMC & GBP Hospital shall not be liable in any manner for any delay occasioned or failure caused in delivering a quotation within the prescribed timeline as aforesaid.
3. The quotationers shall mention their full name, address and contact number on envelopes.
4. Quotations received in due manner are likely to be opened next day after end date of submission, in the Office of the Principal, AGMC, Agartala, unless postponed for any unfrozen reason. Quotationers can be physically present to witness the quotations-opening proceeding (no separate notice will be issued to this effect to any quotationer).
5. Self-attested photocopy of relevant up to dated & valid (i) Trade license (ii) OEM License ,if bidder is not OEM than Authorization certificate from OEM, (IV) PAN Card (VI) GST Registration certificate & Last GST return copy, (Vii) Acknowledgement copy of Income tax Clearance certificate. The Technical (Tender Evaluation) committee of the hospital shall have the right to waive off any of the aforesaid requisites in deserving cases
6. Rate should be quoted inclusive of all taxes.
7. Supply of the equipments should be up to the destination (General Store, AGMC & GBP Hospital, Agartala) at the earliest, but **not later than 30 days after receiving Purchase order.** Penalty @ 1% of the purchase value shall be deducted for every week (or part thereof) of delay in delivery beyond the stipulated date, unless condoned for. In case of supply of lab consumables of sub-standard quality or with any other defects, the supplier shall be liable for compulsorily replacing the same entirely at his own expenses within 15 days of notifying him to this effect.
8. Payment of bills will be made after successful completion of supply as per the terms and conditions of the concerned Purchase Order.
9. The Medical Superintendent & HOD, AGMC& GBP Hospital reserves the right to unilaterally revoke or cancel this SNIQ at his discretion without prior notice; and to reject any quotation or to select a quotation other than the lowest one on justifiable grounds.
10. Approximate requirement is incorporated in the SNIQ which are likely to increase or decrease as per actual requirement. The supply order may be issued once for staggered supply as per requirement.
11. For small instrument/equipment which do not require CAMC, the successful bidders shall have to provide after sales repair service.

Signed by
Sankar Chakraborti

Date: 18-02-2025 13:11:17
Medical Superintendent & HOD
AGMC & GBP Hospital, Agartala

(86)

ANNEXURE-II**Specification for High-throughput semi-automated Urine chemistry Analyzer**

1. The unit should work on reflectance photometer to evaluate the color intensity of each test zone.
2. The unit should have three wavelengths viz 460 nm, 530 nm, 610 nm
3. It should read 12 parameters like Specific gravity, Leukocytes, Nitrite, pH, Protein, Glucose, Ketones, Urobilinogen, Bilirubin, Blood, Microalbumin, Creatinine (automatically calculating ACR)
4. The unit should be a continuous loading system & have through put of around 400 samples/hr
5. The unit should have strip position check system and automatic strip type selection
6. The unit should have concealed waste container for used strips disposal
7. The unit should give results in SI, conventional, arbitrary units and in combination of atleast any two units
8. Company should provide two levels (positive and negative) ready to use liquid stable urine controls
9. Company should provide tri- level grey control strips for instrument QC
10. Instrument should be touch screen with resolution of 800 x 480 pixel.
11. The unit should have built in thermal printing capability
12. Instrument should have option to connect to bar code reader via USB port
13. Unit should have total memory capacity for atleast 2000 measurements including patient samples & controls with details like data and sample ID
14. The analyser should have facility of not allowing sample processing if QC out of range and also facility for QC test timer.
15. Unit should have external ports RS 232, ethernet and also 2 USB ports for LIS connection.
16. Power Supply: 100-240 V; 50-60 HZ.


15/2/25
Professor & Head I/L
Department of Pathology
Agartala Govt. Medical College
Agartala-793006, Tripura